

Calibration Using 15L or 17L Cylinders

ITEMS REQUIRED TO CALIBRATE A STANDARD GAS DETECTOR USING A 15L OR 17L CYLINDER OF GAS:



NUMBER	FUNCTION	PART NUMBER
1	Grey calibration adapter plug with o-ring and two zeroing port nozzles. Fits industry standard internal sensor vent openings. Comes with tubing attached (see 2) Cal Clip for calibrating with splash guard. Comes with tubing attached.	CET-7000-CAP CET-SGC
2	Standard tubing (approx 61 cm / 2 ft long) included and attached to calibration adapter plug	n/a
3	Regulator for fixed flow @0.5 LPM. Fits 34 L, 58 L, 74 L, and 100 L cylinders	REG-0715-CR2
4	Brass fitting to convert fixed flow regulator to attach to 15L and 17 L gas cylinders	REG-ADAPTER
5	Span Gas / Zero Gas cylinder, with air balance or nitrogen balance (electrochemical sensors), with air balance only (catalytic sensors)	various