

ITEMS REQUIRED TO CALIBRATE A STANDARD GAS DETECTOR USING A 34L, 58L, 78L OR 100L CYLINDER OF GAS:



NUMBER	FUNCTION	PART NUMBER
1	Grey calibration adapter plug with o-ring and two zeroing port nozzles. Fits industry standard internal sensor vent openings. Comes with tubing attached (see 2)	CET-7000-CAP
2	Standard tubing (approx 61 cm / 2 ft long) included and attached to calibration adapter plug	n/a
3	Regulator for fixed flow @0.5 LPM. Fits 34 L, 58 L, 74 L, and 100 L cylinders	REG-0715-CR2
4	Span Gas / Zero Gas cylinder, with air balance or nitrogen balance (electrochemical sensors), with air balance only (catalytic sensors)	various