

User Manual

VGW-mini Portable Vacuum Gauge



1. Introduction








VGW-mini is a small and portable professional instrument for vacuum measurement powered by a lithium battery. The 1/4SAE fitting is convenient for vacuum measurement in the HVAC/R field. VGW-mini supports offline data recording, by connecting the meter, users can view data remotely and set vacuum alarms via APP.

1.1 Technical Parameters

| Characteristic | Parameters |
|-----------------------|--|
| Measuring Range | 1-19000microns |
| Vacuum Units | microns、mTorr、inHg、Pa、Torr、KPa、mbar、psia |
| Temperature Units | °C/°F |
| Resolution | 1-400microns 1micron 400-3000microns 10microns 3000-10000microns 100microns 10000-19000microns 250microns |
| Accuracy | 1-10000microns ±10% of Reading ±10microns 10000-19000microns ±20% of Reading |
| Battery | Rechargeable lithium batteries (1000mAh) |
| Operating Temperature | 14~122°F/-10~50°C |
| Storage temperature | 14~140°F/-20~60°C |
| Interface | 1/4 SAE Male Flare |
| Overload | 27.5bar |

1.2 Displays overview



| Number | Element | Function |
|--------|--|--|
| ① | Icon [] | Displays the battery capacity  > 75%  > 50%  > 25%  > 5%  < 5% |
| ② | Icon [] | Display icon when Bluetooth connection |
| ③ | Icon [REC] | Display icon when recording is enabled |
| ④ | Temperature display | <ul style="list-style-type: none"> • Displays the currently measured temperature • Measurement parameter: THzO = evaporation temperature of water Tamb = ambient temperature ΔT = Tamb-THzO • Unit set (°C/°F) |
| ⑤ | Vacuum display | <ul style="list-style-type: none"> • Displays the currently measured vacuum • Unit set (microns、mTorr、inHg、Pa、Torr、KPa、mbar、psia) |

2. Operation Guide

1. Press the power button to turn on the device.

❗ *The instrument displays " - - - - " when ambient pressure is applied to the connections. The display indicates the applied pressure value once the applied pressure is within the measuring range (0 to 19,000 microns).*

2. Search for "**Elitech Tools**" in the APP Store/Google Play or scan the QR code on the back of the product to download the APP with your tablet or smartphone.




3. Open "**Elitech Tools**" and click "**Search nearby Device**". After the device name appears, click the product icon and enter the operation interface after the connection is successful.

❗ *The Bluetooth icon on the product screen is always on when the product is successfully connected to the APP.*

4. Connect the product system and start the test.

2.1 APP Interface Overview





2.2 Unit Setting

Set the unit according to your usage habits.

Click '**microns**' to set vacuum units.

Click '**Settings**' -> '**Temperature Unit**' to set the temperature unit.

2.3 H₂O Status

| | |
|---|----------------------------------|
|  | H ₂ O in liquid state |
|  | H ₂ O in vapor state |




The meter judges the physical state of the water by comparing the ambient temperature and the water saturation temperature corresponding to the vacuum in the system. When the water saturation temperature is less than the ambient temperature ($T_{H_2O} < T_{amb}$), the water is gaseous (evaporated into water vapor), and the moisture in the pipeline can be more effectively removed.

! *Suggestion: If the H₂O state still remain liquid after it reaches the "Evacuation Target", please adjust the "Evacuation Target" and continue pumping until it vaporized.*

2.4 Set Evacuation Target

After setting the Evacuation Target, an alarm notice will be prompted when the vacuum value reaches the set evacuation target.

2.5 Set Decay Target Value for Leak Test

| | | |
|---|--|---|
| Evacuation Target | Decay Target |  |
|  300.0 |  1200.0 | |


Setting the Decay target and conducting a leak test can help you judge the sealing condition of the system!

If you have checked the sealing condition of the system recently (depending on the specific situation), please ignore this step!

Set the appropriate Decay target value according to the actual working conditions, click the TEST button to enter the test time, and click confirm to start the leak test.

2.6 Data Record

① Online Record

APP Online Recording requires the APP to maintain a connection with the device. The recorded data is stored in the phone's memory and does not occupy the storage space of the device. Click the  button on the APP to start recording.

② Offline record

when you need to leave during the recording process, you may turn on the offline recording in the settings, and the device LCD displays the "REC".

! Note: The REC icon flashes when the internal storage space of the device is insufficient.

2.7 Temperature Offline Display Mode



The device can only display two temperature values when used offline.

Through the APP setting, user may switch between TH₂O (saturation temperature) + Tamb (ambient temperature), ΔT (temperature difference) + Tamb (ambient temperature) two display combinations.

3. Precautions

1. Try to avoid using the product close to the heating sources so as not to affect the accuracy of the ambient temperature measurement.
2. When the REC icon is flashing, you need to clear the data in the APP settings for continuous recording.
3. When the product is not used for a long time, it needs to be fully charged and charged once every three months.

4. Product Maintenance



The filter cotton is used in the vacuum chamber to filter impurities and reduce sensor pollution. In order to maintain the best measurement accuracy, normal use needs to be checked every three months (adjust the inspection period according to the frequency of use). If the filter is seriously polluted, it needs to be replaced.

Follow the steps below:

1. Turn the gauge off.
2. Remove the sensor nut with a wrench, and remove the filter cotton.
3. Check whether the filter cotton is contaminated with impurities. Try to wipe it with a paper towel. Replace the filter cotton if it cannot be cleaned .
4. Check whether the O-ring is intact. If damaged, replace it. Before replacement, lubricate the O-ring with vacuum oil .
5. Put the filter cotton into the sensor nut and tighten the nut with a wrench .

If the vacuum sensor inside the cavity is contaminated, follow the methods below to clean it :

1. Inject Inject acetone or alcohol (>70%) into the vacuum cavity with a dropper ora syringe. Tighten the nut and gently shake the gauge.
2. Loosen the nut and drain the fluid from the cavity,repeat such operation 3 to 4 times.
3. Evacuate it or place it for 3 hours until the sensor gets dry.

Elitech[®]
Innovation Preceding All



Authorized Distributor
APL ASIA CO., LTD.

11/129-132 Moo.5 Lamlukka Rd.,T.Kookut, A.Lamlukka, Pathumthani 12130 Thailand.

TEL. 0-2995-4461-3, FAX. 0-2995-4464

www.apl-asia.com

EMAIL : sales@apl-asia.com

LINE@ : @APL-ASIA